



SAFETY AND BUILDINGS DIVISION  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**

**Scott Walker, Governor**  
**Dave Ross, Secretary**

May 9, 2012

WHEATON FRANCISCAN LAB AT ST. FRANCIS  
WHEATON FRANCISCAN HEALTHCARE  
MARK CHARBOGIAN  
3237 S. 16TH ST  
MILWAUKEE WI 53215

BECKMAN-COULTER, INC  
EGON E. KLAVINS  
250 S, KRAEMER BLVD.  
BREA CA 92821

Re: Description: PREFABRICATED PLUMBING SYSTEM  
Manufacturer: BECKMAN-COULTER, INC  
Product Name: DXI 880I CLINICAL CHEMISTRY ANALYZER  
Model Number(s): DXC 880I  
Product File No: 20120169

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of May 2017.

This approval is contingent upon compliance with the following stipulation(s):

- The indirect waste pipe from this product must discharge by means of an air-gap, an air-gap fitting that conforms to ASME A112.1.3 or an air-gap fitting approved by the Department.
- This product has an acceptable built-in air-gap in the water supply to provide protection of the potable water supply.
- These devices must be installed, maintained and serviced as directed by the manufacturer to perform as advertised.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: jerry.thompson@wi.gov